

EN (English)



## Capsule Support Segment

### MODELS & VARIANT

Gupta Segment (9.00mm) Canabrava Gupta Segment (6.00mm)

### DEVICE DESCRIPTION

Capsule support segment (CS Segment) is a modified design of the cionni capsular tension rings. Partial ring made of clear PMMA cover approximately one quadrant.

### INTENDED USE / INDICATION FOR USE

A disposable piece of sterile capsule support segment is used for support of capsule bag during an ophthalmic surgical procedure.

### MATERIALS USED

Polymethyl methacrylate (PMMA)

### TARGET AGE GROUP

Can be used for all age groups.

### INTENDED USER

Any of the registered medical doctor/medical professionals.

### DURATION OF USE

Long term use (more than 30 days).

### STORAGE CONDITIONS

Store at specified temperature and relative humidity (Temperature in between 10°C to 30°C and relative humidity in between 45% to 65%).

### CONTRAINDICATIONS

The device should not be used in patients who show hypersensitivity to any of the materials used in manufacturing of the device.

### HOW SUPPLIED

Capsule Support Segments are supplied sterile, packed in container and then in a peel open pouch. Such pouch is enclosed within an outer printed box. The overall packing contains Capsule Support Segment and medical literature insert, in an outer enclosing box, with external identification stickers.

### CAUTION/WARNING

- Read instructions before use.
- Do not use if package is damaged and consult instructions for use.
- Do not use if the Capsule Support Segments is broken or damaged.
- Do not use the device if the sterilized package is open or damaged.
- The product is designed for single use only.

- The Capsule Support Segments might not have proper spring action.
- The use of this product is restricted to a qualified doctor or medical practitioners.
- Store at specified temperature and humidity.
- Keep out of reach of children.
- Do not re-sterilize the device by any method. Madhu Instruments assumes no liability for the devices that have been re-sterilized.
- Do not use the device after sterility expiry date.

### PRECAUTIONS

- Ensure before use that product/package is not tampered or damaged.
- When handling the product care should be taken to avoid damage from handling. Avoid crushing and crimping damage to the product due to application of surgical instruments such as forceps etc.

### DIRECTION FOR USE

- Check the package sealing integrity to ensure sterility.
- Open the container cap and take out the Capsule Support Segment and wash it with saline water.
- Feed one eyelet of the ring into the capsular bag using a McPherson forceps.
- Then slowly insert the Capsule Support Segment in capsular bag using a lens dilator.
- Insert the second eyelet of the Capsule Support Segment in the bag using a McPherson forceps and check the capsular bag integrity.
- Stabilize the Capsule Support Segment by suturing the fixation arm to the sclera as per approved surgical technique.
- Proceed further to complete the remaining Phaco or IOL inserting procedure.

### EXPIRATION DATE

The expiration date of the product is the sterility expiration date. The product should not be used after the indicated sterility expiration date, the shelf life for sterile products is recommended to 5 years from the date of sterilization.

### RETURN OF DAMAGED PRODUCT

Return the product in its original packing identified by the batch number, purchase information, your reference and reason for return. Please contact your local distributor office regarding product return/exchange.

### REPORTING

Adverse events and/or potentially sight-threatening complications that may

reasonably be regarded as product related and that were not previously expected in nature, severity or incidence must be report to Madhu Instruments Pvt. Ltd.

### DISPOSAL OF THE USED DEVICE

The used device should be disposed off in compliance with applicable local, state and country's laws and regulations.

### EXPLICATION OF SYMBOLS:

Symbol	Meaning
	Caution
	Manufacturer
	Date and Country of Manufacture
	Consult instructions for use
	Keep away from sunlight
	Medical Device
	Do not use hook
	This side up
	Temperature Limit
	Humidity Limitation
	Keep Out of reach of children

Symbol	Meaning
	Maximum stacking
	Handle with care
	Batch Code
	Use by date
	Do not use if package is damaged and consult instructions for use
	Do not re-use
	Single Sterile Barrier System, with inside protective packaging sterilized using Ethylene Oxide
	Single Sterile Barrier System with inside and outside protective packaging, sterilized using Ethylene Oxide
	Do not re-sterilize
	Catalogue Number
	Unique device identifier

### MADHU INSTRUMENTS PVT. LTD.

Mfg. Facility: A 260, Okhla Industrial Area Phase-I, New Delhi-110020 India.  
Regd. Add.: F-90/3D, Okhla Industrial Area Phase-I, New Delhi-110020 India  
Customer Care No.: +91-11-42701030, 42701031  
Email: customercare@madhuinstruments.com  
Website: www.madhuinstruments.com